UNYQ®

ORDERING FORMS

1. ORDER DETAILS

UNYQ Client ID		**Prosthetist**	
Patient/User ref.			
Previous cover (s/n)			
Notes			

2. PRODUCT

Fill in the product details on the corresponding page.

CustomFit	QuickFit	KneeFit	OneFit
FLEX - p. 2 **FIRM** - p. 2	p. 3	p. 4	p. 5

3. MEASURING

Choose the preferred method for taking measurements and images for the selected product. Follow the instructions and fill in the requested measurements.

CustomFit	QuickFit	KneeFit	OneFit
Manual TF - p. 6 **Manual TT** - p. 7 **Scanning TF** - p. 8 **Scanning TT** - p. 9			**Scanning TF** - p. 10

CustomFit

UNYQ

Common details

Prosthesis	Transfemoral	Transtibial
Material	Flex	Firm
Design		
Attachment	Magnets & Fast-clamp	Screws & C-Clamp
Size reduction	0%	5% 10%
Colors	Global	
	Front	Back
	Base	Detail

Details FLEX

Coating	**Glossy**
	Satin
Extras	Engrave

Details FIRM

Coating	**Glossy**
	Matte
Extras	Engrave
	Vinyl
	Knee protector (TF)

QuickFit

UNYQ

Prosthesis	AK/ TF	**Material**	Firm
Design	U	**Coating**	Satin
Prosthesis attachment	C-Clamp (screws)	**Front/Back attachment**	Magnets

Knee

- C-Leg
- 3R80
- Genium
- Kenevo
- Quattro
- Rheo XC
- Allux
- 3R85 Dynion

Size

- S
- L

Colors

- Ö 1
- OB 0
- OB 2
- OB 4
- OB 12
- White
- Grey
- Black
- Blue

Colors

KneeFit

UNYQ

Prosthesis	AK/ TF	**Material**	Firm
Design	U	**Coating**	Satin

Knee

- C-Leg 4
- 3R80
- 3R85 Dynion
- Genium
- Kenevo
- Quattro
- Rheo XC

Colors

- Ö
- OB 0
- OB 2
- OB 4
- OB 12
- White
- Grey
- Black
- Blue

Colors

OneFit

UNYQ

Prosthesis	AK/ TF	Material	Firm

Design

Attachment

Magnets & Fast-clamp ☐ Screws & C-Clamp ☐

Size reduction

0% ☐ 5% ☐ 10% ☐

Colors

Global

Front Back

Base Detail

Coating

Glossy ☐

Matte ☐

Extras

Engrave ☐

Vinyl ☐

TF - Manual Ordering

4 photos sound leg



Take 4 photos of the user wearing the prosthesis, as shown in the images above. The camera must be placed perpendicular to the legs, in a low position, and both legs must be centered on the photo.
Please remember that all prosthetic components must be completely visible, including the socket. Please remove shoes, socks and any object that covers the prosthesis.

4 photos prosthesis



MAX FLEXION

Take a photo of the prosthetic knee from the outside at max flexion.



TOP LATERAL

Position the camera next to the lateral side of the socket looking down.



LATERAL & BACK

Position the camera so the top of the foot shell looks like a straight line.

3 measurements

1 MEASUREMENT
MAX CALF VOLUME



Measure the maximum volume of the calf.

cm

1 MEASUREMENT
LOW CALF VOLUME



Measure the volume at the ankle bone.

cm

1 MEASUREMENT
PROSTHESIS LENGTH



Measure from the bottom edge of the socket to the top edge of the foot cosmesis on the lowest part.

cm

SEND US THE FILES AND FORM

Send us this guide with the order details, measurements and photos to **orders@unyq.com**. If you need further assistance please contact us at orders@unyq.com or +34 854 85 60 92

TT - Manual Ordering

4 photos sound leg



Take 4 photos of the user wearing the prosthesis, as shown in the images above. The camera must be placed perpendicular to the legs, in a low position, and both legs must be centered on the photo. Please remember that all prosthetic components must be completely visible, including the socket. Please remove shoes, socks and any object that covers the prosthesis

3 measurements

MAX CALF VOLUME



Measure the maximum volume of the calf.

cm

LOW CALF VOLUME



Measure the volume at the ankle bone.

cm

PROSTHESIS LENGTH



Measure from the bottom edge of the socket to the top edge of the foot cosmesis on the lowest part.

cm

60 photos of the prosthesis

Please follow the instructions included in the UNYQ BK Kit to set up the prosthesis for the process.

IMPORTANT!

1. Choose a room with good, stable lighting. Avoid opened windows that create glare.
2. Ensure there are no mirrors or reflecting objects
3. Clear the background behind the prosthesis
4. There must be no other people in the room, to avoid interferences.



You must now take a series of photos from your phone or camera as you walk around the prosthesis. You must walk around in 2 different levels ant take 30 photos per level, a total of 60 pictures.

The first series must be taken from chest height looking down, and you must walk around the prosthesis at a slow and steady pace. You must take 30 pictures as you walk around the prosthesis (1 round).

Now you have to take the second series. This may be taken from waist height. You must take again another 30 pictures on this level, as you walk around the prosthesis.

SEND US THE FILES AND FORM

Send us this guide with the order details, measurements and photos to **orders@unyq.com**. If you need further assistance please contact us at orders@unyq.com or +34 854 85 60 92

TF - Scanning Ordering

4 photos sound leg



Take 4 photos of the user wearing the prosthesis, as shown in the images above. The camera must be placed perpendicular to the legs, in a low position, and both legs must be centered on the photo.
Please remember that all prosthetic components must be completely visible, including the socket. Please remove shoes, socks and any object that covers the prosthesis.

3 measurements

MAX CALF VOLUME



Measure the maximum volume of the calf.

cm

LOW CALF VOLUME



Measure the volume at the ankle bone.

cm

PROSTHESIS LENGTH



Measure from the bottom edge of the prosthetic knee to the top edge of the foot cosmesis on its lowest part.

cm

Scanning

NO SHOES

NO SOCKS

.stl
.obj
FORMATS

STRUCTURED LIGHT SCANNERS

- All prosthetic components should be clearly visible on the scan, including the lower part of the socket. Please remove any item that covers the prosthesis (sock, foam, other coverings ...)
- No reflecting ground: Avoid problems with reflective material or dark areas. If your scanner does not receive information from these areas, cover the reflective parts with matt tape, tightly taped to the surface so that it does not create extra bulk.
- We recommend that the environment contrasts in color with the prosthesis.
- Set the scanner to the highest resolution possible.
- Scan the complete prosthesis and the sound leg, with the patient standing up. You can scan both legs in the same file or do it separately.

Recomended scanner: Einstar Scanner; Shining 3D Einscan Pro Series Hand Scanners; Artec Structured Light Hand Scanners; Creaform Hand Scanners; Peel3d 3D Scanner.
Not recommended scanners: 3D Systems / Cubify Sense; Structure Sensor / Core; Microsoft Kinect 3D Scanner; Intel Infrared Sensor Scanners

SEND US THE FILES AND FORM

Send us this guide with the order details, measurements and photos to **orders@unyq.com**. If you need further assistance please contact us at orders@unyq.com or +34 854 85 60 92

TT - Scanning Ordering

4 photos sound leg



Take 4 photos of the user wearing the prosthesis, as shown in the images above. The camera must be placed perpendicular to the legs, in a low position, and both legs must be centered on the photo. Please remember that all prosthetic components must be completely visible, including the socket. Please remove shoes, socks and any object that covers the prosthesis.

4 measurements

MAX CALF VOLUME



Measure the maximum volume of the calf.

cm

LOW CALF VOLUME

Measure the volume at the ankle bone.

cm

TUBE LENGTH



Measure from the bottom edge of the socket to the top edge of the foot cosmesis on its lowest part.

cm

MAX SOCKET VOLUME

Measure the maximum circumference of the socket.

cm

Scanning

NO SHOES

NO SOCKS

.stl
.obj
FORMATS

STRUCTURED LIGHT SCANNERS

- All prosthetic components should be clearly visible on the scan, including the upper part of the socket in transtibial prostheses. Please remove any item that cover the prosthesis (sock, foam, other coverings ...)
- If the patient is wearing a prosthetic knee sleeve, keep it on the prosthesis and fold the remaining fabric into the socket
- No reflecting ground: Avoid problems with reflective material or dark areas. If your scanner does not receive information from these areas, cover the reflective parts with matte tape, tightly taped to the surface so that it does not create extra bulk.
- We recommend that the environment contrasts in color with the prosthesis.
- Set the scanner to the highest resolution possible.
- Scan the complete prosthesis and the sound leg, with the patient standing up. You can scan both legs in the same file or do it separately.

Recomended scanner: Einstar Scanner; Shining 3D Einscan Pro Series Hand Scanners; Artec Structured Light Hand Scanners; Creaform Hand Scanners; Peel3d 3D Scanner.
Not recommended scanners: 3D Systems / Cubify Sense; Structure Sensor / Core; Microsoft Kinect 3D Scanner; Intel Infrared Sensor Scanners

SEND US THE FILES AND FORM

Send us this guide with the order details, measurements and photos to **orders@unyq.com**. If you need further assistance please contact us at orders@unyq.com or +34 854 85 60 92

OneFit - Scanning Ordering

2 photos sound leg



Take 2 photos of the user wearing the prosthesis, as shown in the images above. The camera must be placed perpendicular to the legs, in a low position, and both legs must be centered on the photo. Please remember that all prosthetic components must be completely visible, including the socket. Please remove shoes, socks and any object that covers the prosthesis.

2 measurements



MAX CALF VOLUME

Measure the maximum volume of the calf.

cm



PROSTHESIS LENGTH

Measure from the bottom edge of the prosthetic knee to the top edge of the foot cosmesis on its lowest part.

cm

Scanning

NO SHOES

NO SOCKS

.stl
.obj
FORMATS

STRUCTURED LIGHT SCANNERS



COMPLETE PROSTHESIS



COMPLETE PROSTHESIS SITTING DOWN



SOUND LEG

- All prosthetic components should be clearly visible on the scan, including the lower part of the socket. Please remove any item that covers the prosthesis (sock, foam, other coverings ...)
- No reflecting ground: Avoid problems with reflective material or dark areas. If your scanner does not receive information from these areas, cover the reflective parts with matt tape, tightly taped to the surface so that it does not create extra bulk.
- We recommend that the environment contrasts in color with the prosthesis.
- Set the scanner to the highest resolution possible.
- Scan the complete prosthesis and the sound leg, with the patient standing up. You can scan both legs in the same file or do it separately.

Recomended scanner: Einstar Scanner; Shining 3D Einscan Pro Series Hand Scanners; Artec Structured Light Hand Scanners; Creaform Hand Scanners; Peel3d 3D Scanner.
Not recommended scanners: 3D Systems / Cubify Sense; Structure Sensor / Core; Microsoft Kinect 3D Scanner; Intel Infrared Sensor Scanners

SEND US THE FILES AND FORM

Send us this guide with the order details, measurements and photos to **orders@unyq.com**. If you need further assistance please contact us at orders@unyq.com or +34 854 85 60 92